




IDAHO DEPARTMENT OF HEALTH & WELFARE  
**DIVISION OF PUBLIC HEALTH**

Test Title	EPA 548.1 Endothall
<b>Sample Requirements</b>	<ol style="list-style-type: none"><li>1. Sample type: Water</li><li>2. Minimum sample size: 250 mL</li><li>3. Maximum holding time before analysis: 7 days</li><li>4. Samples must be iced or refrigerated at 4°C until shipment to IBL.</li><li>5. Rejection criteria: Samples may not be analyzed if maximum holding time is exceeded or appropriate temperature is not maintained.</li></ol>
<b>Sampling Materials</b>	<ol style="list-style-type: none"><li>1. Sample container: 1 pre-preserved (Sodium Thiosulfate) 250 mL amber glass bottle. Please obtain sample containers from IBL. When sampling, do not rinse out the preservative.</li></ol> 
<b>Procedural Notes</b>	<ol style="list-style-type: none"><li>1. <a href="#">Drinking Water Analysis - Chemistry Test Request Form</a></li><li>2. Price: \$144</li></ol>
<b>Shipping Instructions</b>	<ol style="list-style-type: none"><li>1. Temperature/preservative instructions: Samples must be shipped in a cooler on ice or frozen gel packs.</li><li>2. Ship to: Idaho Bureau of Laboratories <b>ATTENTION: Chemistry</b> 2220 Old Penitentiary Rd Boise, ID 83712</li></ol>
<b>Reporting and Turnaround Time (TAT)</b>	<ol style="list-style-type: none"><li>1. TAT: 28 days</li><li>2. Results will be mailed unless otherwise requested.</li><li>3. Reference Range:<ol style="list-style-type: none"><li>a. Minimum Detection Limit: 1.38 µg/L</li><li>b. Minimum Reporting Limit: 5 µg/L</li></ol></li></ol>