




IDAHO DEPARTMENT OF HEALTH & WELFARE
DIVISION OF PUBLIC HEALTH

Test Title	EPA 180.1 Determination of Turbidity in Water
Sample Requirements	<ol style="list-style-type: none">1. Sample type: Water2. Minimum sample size: 100 mL3. Maximum holding time before analysis: 48 hours4. Samples must be stored refrigerated until shipment to IBL.5. Rejection criteria: Samples may not be analyzed if maximum holding time is exceeded or appropriate temperature is not maintained.
Sampling Materials	<ol style="list-style-type: none">1. Sample container: Clean plastic container 
Procedural Notes	<ol style="list-style-type: none">1. Drinking Water Analysis - Chemistry Test Request Form2. Price: \$13
Shipping Instructions	<ol style="list-style-type: none">1. Temperature/preservative instructions: Samples must be shipped in a cooler on ice or frozen gel packs.2. Ship to: Idaho Bureau of Laboratories ATTENTION: Chemistry Section 2220 Old Penitentiary Rd Boise, ID 83712
Reporting and Turnaround Time (TAT)	<ol style="list-style-type: none">1. TAT: 5 business days unless also submitted for other tests with longer turnaround times2. Results will be mailed unless otherwise requested.3. Reference Range:<ol style="list-style-type: none">a. Minimum Reporting Limit: 0.1 NTU